



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

PRECISION REPAIR AND CALIBRATION, INC.  
10370 Flanders Street NE  
Blaine, MN 55449  
Brian Downie Phone: 763 762 6250

CALIBRATION

Valid to: May 31, 2025

Certificate Number: 5428.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1, 10</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Calipers <sup>3</sup>	Up to 60 in	$(71 + 33L) \mu\text{in}^9$	Grade 2 gage blocks
Caliper Masters	Up to 12 in	100 $\mu\text{in}$	Grade 2 gage blocks, electronic indicator, surface plate
CMMs <sup>3</sup> – Linear Displacement	Up to 120 in length	$(4 + 4L) \mu\text{in}$	Laser system
Volumetric	Up to 27 in	110 $\mu\text{in}$	Ball bar
Electronic Indicator	Up to 0.05 in	23 $\mu\text{in}^9$	Grade 2 gage blocks
Gage Blocks	Up to 20 in	$(1.6 + 1.9L) \mu\text{in}$	Laser ruler, grade 1 master gage block set
Height Gages <sup>3</sup>	Up to 40 in	$(42 + 8.3L) \mu\text{in}^9$	Grade 2 gage blocks, surface plate

Parameter/Equipment	Range	CMC <sup>2,5</sup> ( $\pm$ )	Comments
Height Masters	Up to 18 in	$(9 + 7.4L) \mu\text{in}^9$	Electronic indicator, gage blocks, surface plate
Indicators <sup>3</sup>	Up to 5 in	$(24 + 29L) \mu\text{in}^9$	Grade 2 gage blocks, micrometer head
Ring Gages – Inside Dimension	Up to 11 in	13 $\mu\text{in}$	Universal measuring machine, master rings
Laser Micrometer <sup>3</sup>	Up to 2 in	14 $\mu\text{in}$	Master pin gages
Length Standards	Up to 36 in	$(34 + 6.2L) \mu\text{in}$	Universal measuring machine, gage blocks, bench micrometer
Micrometers (OD) <sup>3</sup>	Up to 36 in	$(12 + 7.9L) \mu\text{in}^9$	Grade 2 gage blocks
Micrometers (ID) <sup>3</sup>	Up to 40 in	$(34 + 6.2L) \mu\text{in}^9$	Gage blocks, bench micrometer, universal measuring machine
Micrometers – Depth <sup>3</sup>	Up to 12 in	$(520 + 1.1L) \mu\text{in}$	Mike master, gage blocks
Micrometer Heads	Up to 2 in	44 $\mu\text{in}^9$	Gage blocks, electronic indicator, optical flat, universal measuring machine
Mike Masters	Up to 6 in	61 $\mu\text{in}$	Gage blocks, surface plate, electronic indicator
Pin Gages (OD) <sup>3</sup>	Up to 1 in	15 $\mu\text{in}$	Master pin gage, laser micrometer
Plug Gages (OD) <sup>3</sup>	Up to 20 in	$(16 + 1.1L) \mu\text{in}$	Universal measuring machine, grade 2 gage blocks

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
Bore Gages <sup>3</sup>	Up to 12 in	(210 + 19L) μin	Gage blocks, gage block accessory, master ring gages
Optical Comparators <sup>3</sup>	Up to 12 in Linearity Magnification	100 μin 53 μin	Glass scales, precision balls
Microscopes <sup>3</sup>	Up to 12 in Linearity Magnification	100 μin 53 μin	Gage blocks, glass scales
VMMs <sup>3</sup>	Up to 12 in Linearity Magnification	110 μin 180 μin	Glass scales, laser
Master Setting Disc	Up to 20 in	(16 + 1.1L) μin	Gage blocks, universal measuring machine
Precision Balls	Up to 2 in	22 μin	Universal measuring machine
Snap Gages <sup>3</sup>	Up to 4 in	160 μin	Gage blocks
Surface Plates – Overall Flatness <sup>3</sup>	Up to 16 ft diagonal	57 μin	Precision levels
Local Area Flatness <sup>3</sup> (repeat reading)	Up to 16 ft diagonal	22 μin	Repeat-o-meter
Thread Plug Gages – Thread Set Plugs <sup>3</sup>	Up to 10 in	85 μin	Bench micrometer, grade 2 gage blocks, thread measuring wires
Thread Ring Gages	Pitch Diameter: Up to 10 in	100 μin	Thread set plug masters
	Minor Diameter: Up to 0.500 in	110 μin	VMM
	(0.500 to 10.00) in	51 μin	ULM

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
Tapered Thread Plug Gages	Up to 7 in	520 μin Basic 190 μin PD	Thread measuring wires, bench micrometer, O.D. micrometer
Tapered Thread Ring Gages, Crest Check	Up to 7 in	520 μin Basic 520 μin Standoff	Master tapered thread plugs, O.D. micrometer
Rulers	Up to 72 in	(1500 + 1.8L) μin	Microscope, gage blocks
Rulers <sup>3</sup>	Up to 72 in	(880 + 8.5L) μin	Gage blocks
Tape Measures <sup>3</sup>	Up to 600 in	4800 μin <sup>9</sup>	Gage blocks
Angle Blocks	Up to 45°	64 μin	Gage blocks, sine bar, indicator, surface plate
Angle Leaf	Up to 12 in	(250 + 1.7L) μin	VMM
1-2-3 Blocks	Up to 6 in	35 μin	Gage blocks, surface plate, indicator
Radius Gages	Up to 12 in	(250 + 1.7L) μin	VMM
V-Blocks	Up to 6 in	110 μin	Pin gages, surface plate, indicator
Parallels & Straight Edges <sup>3</sup>	Up to 48 in	57 μin	Gage blocks, indicator surface plate
Steel Squares	Up to 18 in	(26 + 6.5L) μin	Surface plate, indicator, master square
Feeler Gages	Up to 0.5 in	93 μin	Gage blocks, super micrometer

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Comptor Gage Indicators	Up to 0.04 in	200 μin	Comptor master
Profilometers <sup>3</sup>	Up to 250 μin	6.4 μin	Profilometer standard
Protractors	(0 to 360)°	0.12°	Angle blocks, surface plate, master square
Chamfer Gages <sup>3</sup>	Up to 3 in	99 μin <sup>9</sup>	Chamfer checker gage

## II. Dimensional Testing/Calibration<sup>8</sup>

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
Dimensional Measurement	Up to 6 in Up to 20 in Up to 18 in Up to 1 in Up to 12 in	(71 + 33L) μin <sup>9</sup> 44 μin (9.1 + 7.4L) μin (12 + 7.9L) μin <sup>9</sup> (250 + 1.7L) μin	Caliper Bench micrometer Indicator Micrometer VMM

## III. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
DC Voltage – Generate <sup>3</sup>	Up to 330 mV 330 mV to 3.3 V (3.3 to 33) V (33 to 330) V 100 V to 1 kV	1 μV + 16 μV/V 6 μV + 9 μV/V 60 μV + 10 μV/V 0.59 mV + 14 μV/V 5.9 mV + 14 μV/V	Fluke 5522A multi product calibrator

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments	
DC Voltage – Measure <sup>3</sup>	Up to 100 mV 100 mV to 1 V (1 to 10) V (10 to 100) V 100 V to 1 kV  (1 to 6) kV  (6 to 20) kV (20 to 35) kV @ (20 to 30) °C  (35 to 40) kV	0.64 µV 3.9 µV 31 µV 0.55 mV 5.3 mV  1.1 mV/V + 2.3 V  2.3 mV/V + 220 mV 1.2 mV/V + 160 mV  2.3 mV/V + 67 mV	HP 3458A Opt 002 multimeter    Fluke 187 w/ Fluke 80K6 probe  Fluke 187 w/ Fluke 80K40 probe	
DC Current – Generate <sup>3</sup>	Up to 330 µA 330 µA to 3.3 mA (3.3 to 33) mA (33 to 330) mA 330 mA to 1.1 A (1.1 to 3) A (3 to 11) A (11 to 20) A	20 nA + 0.12 mA/A 0.25 µA + 78 µA/A 0.61 µA + 78 µA/A 6.1 µA + 78 µA/A 66 µA + 0.16 mA/A 66 µA + 0.3 mA/A 0.70 mA + 0.39 mA/A 0.82 mA + 0.78 mA/A	Fluke 5522A multi product calibrator	
DC Current – Measure <sup>3</sup>	Up to 100 nA 100 nA to 1 µA (1 to 10) µA (10 to 100) µA 100 µA to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A  (1 to 1200) A	40 pA + 35 µA/A 40 pA + 25 µA/A 0.1 nA + 25 µA/A 0.8 nA + 25 µA/A 5 nA + 25 µA/A 50 nA + 25 µA/A 0.5 µA + 40 µA/A 10 µA + 0.12 mA/A  2.3 mA + 2.5 mA/A	HP 3458A opt 002 multimeter       HP 3458A multimeter w/ current shunts Empro B-1200-100	
DC Power – Generate <sup>3</sup>	33 mV to 1.02 kV: (0.33 to 329.99) mA (0.33 to 2.9999) A (3 to 20.5) A	0.01 µW to 337 W (0.01 to 3060) W (3.06 to 20.91) kW	5.8 µW + 240 µW/W 4 µW + 240 µW/W 5.4 µW + 690 µW/W	Fluke 5522A multi product calibrator

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Capacitance <sup>3</sup> – Generate	220 pF to 399.9 pF (0.4 to 1.0999) nF (1.1 to 3.2999) nF (3.3 to 10.9999) nF (11 to 32.9999) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.9999) μF (11 to 32.9999) μF (33 to 109.999) μF (110 to 329.999) μF (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.9999) mF (11 to 32.9999) mF (33 to 110) mF	0.43 % nominal + 12 pF 0.44 % nominal + 12 pF 0.45 % nominal + 12 pF 0.22 % nominal + 12 pF 0.23 % nominal + 12 pF 0.23 % nominal + 12 pF 0.26 % nominal + 360 pF 0.23 % nominal + 1.2 nF 0.28 % nominal + 2.9 nF 0.25 % nominal + 11 nF 0.43 % nominal + 28 nF 0.44 % nominal + 110 nF 0.45 % nominal + 320 nF 0.39 % nominal + 1.3 μF 0.54 % nominal + 1.9 μF 0.52 % nominal + 9.9 μF 0.87 % nominal + 36 μF 1.2 % nominal + 94 μF	Fluke 5522A multi product calibrator
Resistance <sup>3</sup> – Generate	Up to 10.9999 Ω (11 to 32.9999) Ω (33 to 109.9999) Ω (110 to 329.9999) Ω (330 to 1099.999) Ω (1.1 to 3.299 999) kΩ (3.3 to 10.999 99) kΩ (11 to 32.999 99) kΩ (33 to 109.9999) kΩ (110 to 329.9999) kΩ (.330 to 1.099 999) MΩ (1.1 to 3.299 999) MΩ (3.3 to 10.999 99) MΩ (11 to 32.999 99) MΩ (33 to 109.9999) MΩ (110 to 329.9999) MΩ (330 to 1100) MΩ	110 μΩ/Ω + 140 μΩ 30 μΩ/Ω + 63 μΩ 25 μΩ/Ω + 110 μΩ 26 μΩ/Ω + 38μΩ 26 μΩ/Ω + 7.7 μΩ 25 μΩ/Ω + 58μΩ 84 μΩ/Ω + 5.8mΩ 2.6 mΩ/Ω + 5.8mΩ 1.0 mΩ/Ω + 58mΩ 3 mΩ/Ω + 81 mΩ 33 mΩ/Ω + 830 mΩ 36 mΩ/Ω + 810 mΩ 41 mΩ/Ω + 810 mΩ 480 mΩ/Ω + 5.8 Ω 560 mΩ/Ω + 12 Ω 780 mΩ/Ω + 12 Ω 1.4 Ω/Ω + 57 Ω	Fluke 5522A multi product calibrator
Resistance <sup>3</sup> – Measure	Up to 10 Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ (1 to 10) MΩ (10 to 100) MΩ 100 MΩ to 1 GΩ	58 μΩ + 10 μΩ/Ω 58 μΩ + 12 μΩ/Ω 0.58 mΩ + 10 μΩ/Ω 5.8 mΩ + 10 μΩ/Ω 57 mΩ + 10 μΩ/Ω 0.58 Ω + 10 μΩ/Ω 5.8 Ω + 50 μΩ/Ω 58 mΩ + 0.5 mΩ/Ω 0.58 Ω + 5 mΩ/Ω	HP 3458A opt 002 multimeter

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Simulation of Thermocouple Indicators <sup>3</sup>			
Type B	(600 to 800) °C (800 to 1000) °C (1000 to 1550) °C (1550 to 1820) °C	0.42 °C 0.35 °C 0.31 °C 0.27 °C	Fluke 5522A-600 multi product calibrator
Type C	(0 to 150) °C (150 to 600) °C (650 to 1000) °C (1000 to 1800) °C (1800 to 2316) °C	0.24 °C 0.20 °C 0.24 °C 0.39 °C 0.64 °C	
Type E	(-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.38 °C 0.13 °C 0.11 °C 0.13 °C 0.17 °C	
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.21 °C 0.13 °C 0.11 °C 0.14 °C 0.19 °C	
Type K	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C	0.26 °C 0.15 °C 0.13 °C 0.20 °C 0.31 °C	
Type L	(-200 to -100) °C (-100 to 800) °C (800 to 900) °C	0.37 °C 0.27 °C 0.18 °C	
Type N	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C	0.31 °C 0.18 °C 0.16 °C 0.16 °C 0.23 °C	
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.48 °C 0.29 °C 0.27 °C 0.31 °C	



Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Simulation of Thermocouple Indicators <sup>3</sup> (cont)			
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1767) °C	0.48 °C 0.31 °C 0.29 °C 0.35 °C	Fluke 5522A-600 multi product calibrator
Type T	(-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.48 °C 0.19 °C 0.13 °C 0.11 °C	
Type U	(-200 to 0) °C (0 to 600) °C	0.56 °C 0.28 °C	
Electrical Simulation of RTD Indicators <sup>3</sup> –			
Pt 385, 100 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.04 °C 0.05 °C 0.07 °C 0.07 °C 0.09 °C 0.09 °C 0.21 °C	Fluke 5522A multi product calibrator
Pt 3926, 100 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C	0.04 °C 0.05 °C 0.07 °C 0.08 °C 0.09 °C 0.10 °C	
Pt 3916, 100 Ω	(-200 to -190) °C (-190 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.25 °C 0.04 °C 0.05 °C 0.06 °C 0.06 °C 0.07 °C 0.08 °C 0.08 °C 0.21 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Simulation of RTD Indicators <sup>3</sup> – (cont)			
Pt 385, 200 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.03 °C 0.03 °C 0.04 °C 0.04 °C 0.11 °C 0.12 °C 0.12 °C 0.14 °C	Fluke 5522A multi product calibrator
Pt 385, 500 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.03 °C 0.04 °C 0.05 °C 0.06 °C 0.07 °C 0.07 °C 0.08 °C 0.09 °C	
Pt 385, 1000 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.03 °C 0.03 °C 0.03 °C 0.05 °C 0.06 °C 0.06 °C 0.07 °C 0.22 °C	
PtNi 385, 120 Ω	(-80 to 0) °C (0 to 100) °C (100 to 260) °C	0.06 °C 0.07 °C 0.13 °C	
Cu 427, 10 Ω	(-100 to 260) °C	0.30 °C	
AC Voltage – Generate <sup>3</sup>			
(1 to 33) mV	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	7.4 μV + 0.62 mV/V 7.4 μV + 0.12 mV/V 7.4 μV + 0.16 mV/V 7.4 μV + 0.78 mV/V 10 μV + 2.7 mV/V 39 μV + 6.2 mV/V	Fluke 5522A multi product calibrator

Parameter/Range	Frequency	CMC <sup>2, 4</sup> (±)	Comments
AC Voltage – Generate <sup>3</sup> (cont)			
(33 to 330) mV	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	8.5 μV + 0.39 mV/V 8.5 μV + 0.11 mV/V 8.5 μV + 0.13 mV/V 8.5 μV + 0.27 mV/V 26 μV + 0.62 mV/V 55 μV + 1.5 mV/V	Fluke 5522A multi product calibrator
330 mV to 3.3 V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	3.4 mV + 0.23 mV/V 4.1 mV + 0.12 mV/V 0.18 mV + 0.15 mV/V 0.18 mV + 0.23 mV/V 0.2 mV + 0.55 mV/V 0.5 mV + 1.9 mV/V	
(3.3 to 33) V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.77 mV + 0.23 mV/V 0.74 mV + 0.12 mV/V 0.74 mV + 0.19 mV/V 0.74 mV + 0.27 mV/V 1.4 mV + 0.7 mV/V	
(33 to 330) V	45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	6 mV + 0.15 mV/V 7.4 mV + 0.16 mV/V 7.4 mV + 0.2 mV/V 7.4 mV + 0.23 mV/V 39 mV + 1.6 mV/V	
330 V to 1 kV	45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	58 mV + 0.23 mV/V 58 μV + 0.2 mV/V 59 mV + 0.23 mV/V	
AC Voltage – Measure <sup>3</sup>			
(1 to 10) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 250) kHz	350 μV/V + 3.5 μV 230 μV/V + 13 μV 350 μV/V + 13 μV 1.2 mV/V + 13 μV 5.8 mV/V + 13 μV 46 mV/V + 23 μV	HP 3458A Opt 002 multimeter

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Measure <sup>3</sup> (cont)			
(10 to 100) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	4.1 μV/V + 180 μV 7.5 μV/V + 31 μV 16 μV/V + 29 μV 34 μV/V + 28 μV 92 μV/V + 25 μV 3.5 mV/V + 120 μV 1.2 mV/V + 120 μV 1.7 mV/V + 40 μV	HP 3458A opt 002 multimeter
100 mV to 1 V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	4 μV/V + 180 μV 7.9 μV/V + 25 μV 160 μV/V + 26 μV 340 μV/V + 26 μV 920 μV/V + 25 μV 3.5 mV/V + 120 μV 12 mV/V + 140 μV 17 mV/V + 400 μV	
(1 to 10) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	79 μV/V + 500 μV 79 μV/V + 260 μV 160 μV/V + 260 μV 340 μV/V + 270 μV 920 μV/V + 250 μV 4.1 mV/V + 180 μV 120 mV/V + 1.5 mV 120 mV/V + 1.5 mV	
(10 to 100) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	230 μV/V + 4.8 mV 230 μV/V + 2.5 mV 230 μV/V + 2.6 mV 400 μV/V + 2.7 mV 1.4 mV/V + 2.6 mV 4.6 mV/V + 12 mV 17 mV/V + 12 mV	
100 V to 1 kV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	460 μV/V + 47 mV 460 μV/V + 24 mV 690 μV/V + 24 mV 1.4 mV/V + 24 mV 3.5 mV/V + 23 mV	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Measure <sup>3</sup> (cont)			
(1 to 6) kV	DC to 1 kHz	2.6 V + 4.3 mV/V	Fluke 187 w/ Fluke 80K6 probe
(6 to 40) kV	60 Hz	58 V + 4.6 mV/V	Fluke 187 w/ Fluke 80K40 probe
AC Current – Generate <sup>3</sup>			
(29 to 329.99) µA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.12 µA + 1.9 nA/A 0.12 µA + 1.4 nA/A 0.12 µA + 1.2 nA/A 0.17 µA + 2.9 nA/A 0.23 µA + 6.9 nA/A 0.46 µA + 14 nA/A	Fluke 5522A multi product calibrator
(0.33 to 3.299 99) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	170 nA + 1.8 µA/A 170 nA + 1.2 µA/A 170 nA + 920 nA/A 230 nA + 1.8 µA/A 350 nA + 4.6 µA/A 690 nA + 9.2 µA/A	
(3.3 to 32.9999) mA	(10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	2.3 µA + 1.7 µA/A 2.3 µA + 420 nA/A 2.3 µA + 760 nA/A 3.5 µA + 1.9 µA/A 4.6 µA + 3.7 µA/A	
(33 to 329.999) mA	(10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	23 µA + 1.7 µA/A 23 µA + 420 nA/A 57 µA + 950 nA/A 110 µA + 2 µA/A 230 µA + 3.8 µA/A	
(0.33 to 2.999 99) A	(10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	110 µA + 1.8 mA/A 91 µA + 660 µA/A 110 mA + 5.9 mA/A 5.8 mA + 23 mA/A	
(3 to 10.9999) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	2.4 mA + 580 µA/A 2.4 mA + 920 µA/A 2.4 mA + 29 mA/A	
(11 to 20.5) A	45 Hz to 1 kHz (1 to 5) kHz	5.7 mA + 1.5 mA/A 5.8 mA + 29 mA/A	



Parameter/Equipment	Range	CMC <sup>2,4</sup> ( $\pm$ )	Comments
AC Power – Generate <sup>3</sup> (33 to 329.99) mV (3.3 to 329.99) mA (0.33 to 2.1999) A (2.2 to 20.5) A	0.1 $\mu$ W to 108 mW 109 mW to 725 mW 726 mW to 6.76 W	180 $\mu$ W + 1.5 mW/W 160 $\mu$ W + 8.6 mW/W 160 $\mu$ W + 1.5 mW/W	Fluke 5522A multi product calibrator
AC Power – Generate <sup>3</sup> (0.33 to 1020) V (3.3 to 329.99) mA (0.33 to 2.1999) A (2.2 to 20.5) A	1 mW to 336 W 110 mW to 2243 W 726 mW to 20 910 W	180 $\mu$ W + 1.4 mW/W 84 $\mu$ W + 1.4 mW/W 160 $\mu$ W + 1.4 mW/W	Fluke 5522A multi product calibrator
Oscilloscopes <sup>3</sup> –  DC Signal Into 50 $\Omega$ Into 1 M $\Omega$  Square Wave Signal Into 50 $\Omega$  Into 1 M $\Omega$  Leveled Sine Wave Amplitude 5 mV to 5.5 V  Leveled Sine Wave Flatness (relative to 50 kHz) 5 mV to 5.5 V  Time Marker 50 $\Omega$ Source & Period  Rise Time	  0 mV to 6.6 V 0 mV to 130 V  1 mV to 6.6 V <sub>p-p</sub> 10 Hz to 10 kHz 1 mV to 130 V <sub>p-p</sub> 10 Hz to 10 kHz  50 kHz reference 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz  50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz  5 s to 50 ms 20 ms to 2 ns  (200 to 350) ps 1 kHz to 10 MHz	  200 $\mu$ V/V + 2 mV 330 $\mu$ V/V + 2 mV  2 mV/V + 30 $\mu$ V 0.76 mV/V + 30 mV  16 mV/V + 0.23 mV 27 mV/V + 0.23mV 31 mV/V + 0.23 mV 47 mV/V + 0.23 mV  12 mV/V + 78 $\mu$ V 16 mV/V + 78 $\mu$ V 31 mV/V + 78 $\mu$ V  19 $\mu$ s/s + 54 $\mu$ s 1.9 $\mu$ s/s  78 ps	Fluke 5522A SC600 multi product calibrator

IV. Electrical – RF/Microwave

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
RF Power – Measure			
(20 to 30) dBm	30 MHz to 2 GHz (2 to 4.2) GHz (4.2 to 18) GHz (18 to 26.5) GHz	0.33 dB 0.34 dB 0.51 dB 0.43 dB	N5531X Measuring receiver (includes N9030B, N9091EMOE, N1913A, N5532B)
(0 to 20) dBm	30 MHz to 2 GHz (2 to 4.2) GHz (4.2 to 18) GHz (18 to 26.5) GHz	0.26 dB 0.28 dB 0.47 dB 0.38 dB	
(-5 to 0) dBm	30 MHz to 2 GHz (2 to 4.2) GHz (4.2 to 18) GHz (18 to 26.5) GHz	0.26 dB 0.28 dB 0.47 dB 0.38 dB	
(-10 to -5) dBm	30 MHz to 2 GHz (2 to 4.2) GHz (4.2 to 18) GHz (18 to 26.5) GHz	0.27 dB 0.29 dB 0.48 dB 0.39 dB	
(-136 to 20) dBm	100 Hz to 26.5 GHz	0.67 dB	
RF Power – Generate			96270A Leveled microwave output
(-100 to +24) dB	1 mHz to 4 GHz	0.70 dB	
(-100 to +20) dB	(1.4 to 20) GHz	1.3 dB	
(-100 to +18) dB	(20 to 26.5) GHz	1.7 dB	



Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Level Sine Wave – Generate			
(-130 to -94) dBm	(10 to 128) MHz 128 MHz to 3 GHz	0.87 dB 1.7 dB	96270A w/ 96040A-50
(-94 to -84) dBm	100 kHz to 10 MHz (10 to 128) MHz (128 to 300) MHz 300 MHz to 3 GHz	0.67 dB 0.47 dB 0.67 dB 1.2 dB	
(-84 to -74) dBm	100 kHz to 10 MHz (10 to 128) MHz (128 to 300) MHz 300 MHz to 1.4 GHz (1.4 to 4) GHz	0.67 dB 0.27 dB 0.47 dB 0.67 dB 1.2 dB	
(-74 to -48) dBm	100 kHz to 100 MHz (10 to 300) MHz 300 MHz to 1.4 GHz (1.4 to 4) GHz	0.37 dB 0.27 dB 0.57 dB 0.67 dB	
(-48 to -17) dBm	10 Hz to 100 kHz 100 kHz to 128 MHz (128 to 300) MHz 300 MHz to 1.4 GHz (1.4 to 3) GHz (3 to 4) GHz	0.20 dB 0.22 dB 0.24 dB 0.37 dB 0.47 dB 0.67 dB	
(-17 to +14) dBm	10 Hz to 100 kHz 100 kHz to 128 MHz (128 to 300) MHz 300 MHz to 1.4 GHz (1.4 to 4) GHz	0.20 dB 0.22 dB 0.24 dB 0.37 dB 0.47 dB	
(+14 to +20) dBm	10 Hz to 100 kHz 100 kHz to 128 MHz (128 to 300) MHz 300 MHz to 1.4 GHz	0.20 dB 0.22 dB 0.24 dB 0.37 dB	
(+20 to +24) dBm	10 Hz to 100 kHz 100 kHz to 128 MHz	0.20 dB 0.22 dB	
Level Sine Wave – Microwave Output			
(-100 to +24) dBm	10 Hz to 4 GHz (4 to 26.5) GHz	0.5 dB 1.0 dB	96270A Phase noise reference source

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Attenuation – Measure  10 Hz to 128 MHz	(0 to 55) dB (55 to 64) dB (64 to 74) dB (74 to 100) dB (100 to 116) dB	0.08 dB 0.09 dB 0.11 dB 0.13 dB 0.21 dB	96270A w/ N9030A & NRP-Z55.03
Amplitude Modulation – Measure  (Carrier: 2 Hz to 26.5 GHz)  (Carrier: 10 MHz to 26.5 GHz; Rate: 38 Hz to 243 kHz)	Depth: (5 to 99) %  Depth: (5 to 99) %	2.0 %  2.0 %	N9030B PXA Signal analyzer
Amplitude Modulation – Generate  (Carrier: 50 kHz to 4 GHz; Rate: 1 Hz to 100 kHz)	Depth: (5 to 10) %  Depth: ( 10 to 99) %	3.3 %  3.9 %	96270A with 96040A-50
Frequency Modulation – Measure  (Carrier: 2 Hz to 26.5 GHz; Rate: 1 Hz to 12.5 MHz)  (Carrier: 10 MHz to 26.5 GHz; Rate: 10 Hz to 1 MHz)	Deviation: Up to 10 MHz  Deviation: Up to 10 MHz	300 mHz/Hz + 1.4 Hz  300 mHz/Hz + 1.4 Hz	N5531X Measuring receiver
Frequency Modulation – Generate  (Carrier: 9 MHz to 31.25 MHz; Rate: 1 Hz to 50 kHz)	Deviation: 10 Hz to 300 kHz	130 mHz/Hz	96270A w/ 96040A-50

Parameter/Equipment	Range	CMC <sup>2,5</sup> ( $\pm$ )	Comments
Frequency Modulation – Generate (cont)  (Carrier: 31.25 MHz to 125 MHz; Rate: 50 kHz to 300 kHz)	Deviation: 10 Hz to 300 kHz	310 mHz/Hz	96270A w/96040A- 50
Frequency Modulation – Generate (cont)  (Carrier: 31.25 MHz to 125 MHz; Rate: 1 Hz to 50 kHz)  (Carrier: 31.25 MHz to 125 MHz; Rate: 50 kHz to 300 kHz)  (Carrier: 125 MHz to 4 GHz; Rate: 1 Hz to 50 kHz)  (Carrier: 125 MHz to 4 GHz; Rate: 50 kHz to 300 kHz)	Deviation: 10 Hz to 750 kHz  Deviation: 10 Hz to 750 kHz  Deviation: 10 Hz to 4.8 MHz  Deviation: 10 Hz to 4.8 MHz	130 mHz/Hz  310 mHz/Hz  130 mHz/Hz  310 mHz/Hz	96270A w/96040A- 50
Phase Modulation – Measure  (Carrier: 2 Hz to 26.5 GHz; Rate: 50 Hz to 100 kHz)  (Carrier: 10 MHz to 26.5 GHz; Rate: 100 kHz to 1 MHz)	Deviation: (0.01 to 25 000) rad  Deviation: (0.01 to 25 000) rad	7.9 x 10 <sup>-3</sup> rad/rad  7.9 x 10 <sup>-3</sup> rad/rad	N5531X Measuring receiver

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
Phase Modulation – Generate  (Carrier: 9 MHz to 4 GHz; Rate: 1 Hz to 50 kHz)  (Carrier: 9 MHz to 4 GHz; Rate: 1 Hz to 50 kHz)	Deviation: Up to 1000 rad  Deviation: Up to 1000 rad	3.0 x 10 <sup>-3</sup> rad/rad  3.0 x 10 <sup>-3</sup> rad/rad	96270A Phase noise reference source 41

#### V. Mechanical

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
Force Gage <sup>3</sup> (Tension & Compression)	Up to 10 lb (10 to 110) lb (110 to 500) lb	0.05 lb 0.15 lb 0.46 lb	Test weights
Laboratory Balance <sup>3</sup>	Up to 6000 g	(0.01 + 0.000 000 14 <i>W</i> ) mg	Weight kit
Industrial Scales <sup>3</sup>	Up to 500 lb	0.012 lb <sup>9</sup>	Weight kit
Pipettes <sup>3</sup>	(0.5 to 100) µL (100 to 10 000) µL	(0.0064 + 0.025 <i>vol</i> ) µL (1.4 + 0.0002 <i>vol</i> ) µL	Laboratory balances
Pressure Gages <sup>3</sup>	(10 to 10 000) psig  (-15 to 0.001) psig (0.001 to 300) psig (300 to 10 000) psig	0.035 %  0.042 % 0.042 % 0.035 %	Dead weight tester  Pressure monitor

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
Rockwell Hardness Testers <sup>3</sup>	HRC: High Medium Low  HRBW: High Medium Low  HR15N: High Medium Low  HR30N: High Medium Low  HR45N: High Medium Low  HR15T: High Medium Low  HR30T: High Medium Low	1.3 HRC 1.2 HRC 1.2 HRC  1.5 HRBW 1.6 HRBW 2.3 HRBW  1.2 HR15N 1.2 HR15N 1 HR15N  1.2 HR30N 1.3 HR30N 1.5 HR30N  1.4 HR45N 1.7 HR45N 1.7 HR45N  1.1 HR15T 1.1 HR15T 1.3 HR15T  1.1 HR30T 1.1 HR30T 1.4 HR30T	Test blocks
Brinell Hardness Testers <sup>3</sup>	(100 to 240) HBW (240 to 600) HBW Above 600 HBW	1 HBW 1.9 HBW 5.1 HBW	Test blocks
Vickers Hardness Testers <sup>3</sup>	(170 to 700) HV	15 HV	Test blocks
Knoop Hardness Testers <sup>3</sup>	(170 to 200) HK (200 to 400) HK (400 to 700) HK	1 HK 2.4 HK 5.5 HK	Test blocks

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
Torque Transducers Torque Watches	Up to 10 ozf·in (10 to 50) ozf·in 50 ozf·in to 10 lbf·in (10 to 50) lbf·in (50 to 100) lbf·in (100 to 250) lbf·in (250 to 600) lbf·in	0.023 ozf·in 0.05 ozf·in 0.005 lbf·in 0.013 lbf·in 0.053 lbf·in 0.08 lbf·in 0.16 lbf·in	Dead weights, torque arms & wheels
Torque Wrenches <sup>3</sup>	Up to 26 ozf·in 26 ozf·in to 50 lbf·in (50 to 500) lbf·in 500 lbf·in to 150 lbf·ft (150 to 250) lbf·ft (250 to 600) lbf·ft (600 to 1000) lbf·ft	0.23 ozf·in 0.23 lbf·in 0.94 lbf·in 0.38 lbf·ft 0.40 lbf·ft 3.5 lbf·ft 3.6 lbf·ft	Torque transducer system

#### VI. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2,7</sup> (±)	Comments
Temperature – Measure	(-180 to 0) °C (0 to 100) °C (100 to 200) °C (200 to 400) °C (400 to 600) °C	0.000 03 °C/°C + 0.026 °C 0.000 065 °C/°C + 0.025 °C 0.000 14 °C/°C + 0.018 °C 0.000 11 °C/°C + 0.025 °C 0.000 051 °C/°C + 0.047 °C	Fluke 1524 thermometer, Fluke 5615/6, Fluke 5609
Temperature – Measuring Equipment	(-30 to 140) °C (-30 to 670) °C	0.000 072 °C/°C + 0.022 °C 0.000 066 °C/°C + 0.021 °C	Fluke 1523A, Fluke 5615/6, Fluke 5609, Fluke 9011 Dry well
Humidity <sup>3</sup> – Measure	(15 to 90) % RH (90 to 95) % RH	1.4 % RH 2.2 % RH	Vaisala HMI41, HMP46 humidity indicator & probe

Parameter/Equipment	Range	CMC <sup>2,7</sup> (±)	Comments
Humidity <sup>3</sup> – Measuring Equipment	(15 to 90) % RH (90 to 95) % RH	1.5 % RH 1.7 % R H	Vaisala HMI41, HMP46 humidity indicator & probe, Test Equity 123H temperature/humidity chamber
	(10 to 85) °C	0.032 °C	Fluke 1523A, Fluke 5615/6

## VII. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2,7</sup> (±)	Comments
Frequency – Generate <sup>3</sup>	DC to 600 MHz	0.061 µHz/Hz + .041 Hz	Fluke 5522A multi product calibrator w/ Fluke PM6681R
	600 MHz to 40 GHz	0.77 µHz/Hz + 1.3 Hz	Agilent E8257D signal generator w/ Agilent 53152A
Frequency – Measure <sup>3</sup>	0.11 Hz to 1.3 GHz	0.000 44 µHz/Hz	Fluke PM6681R
	10 Hz to 46 GHz	13 µHz/Hz	Agilent 53152A
	DC to 12.4 GHz	4.7 µHz/Hz	Agilent 53131A frequency counters
Frequency – Dissemination	10 MHz	5.8 x 10 <sup>-10</sup> Hz	Fluke PM6681R
Timers & Stopwatches <sup>3</sup>	(0 to 19.99) sec/day	0.061 sec/day	Timometer 4500

<sup>1</sup> This laboratory offers commercial dimensional testing, calibration and field calibration service.

- <sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- <sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- <sup>4</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percentage or fraction of the reading plus a fixed floor specification.
- <sup>5</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in inches.  $W$  is the applied weight in milligrams. In the statement of CMC, the value is defined as the percentage of reading, unless otherwise noted. In the statement of CMC,  $vol$  is defined as the volume of the reading.
- <sup>6</sup> In the statement of CMC, percentages are percentage of reading, unless otherwise indicated.
- <sup>7</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.
- <sup>8</sup> This laboratory meets *R205 – Specific Requirements: Calibration Laboratory Accreditation Program* for the types of dimensional tests listed above and is considered equivalent to that of a calibration.
- <sup>9</sup> In the statement of CMC the resolution of the instrument is not taken into account. Addition of 0.6 times the resolution will be added to the uncertainty for the individual unit under test.
- <sup>10</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.





# Accredited Laboratory

A2LA has accredited

## PRECISION REPAIR AND CALIBRATION, INC.

*Blaine, MN*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

Presented this 21<sup>st</sup> day of November 2023.

A handwritten signature in blue ink, appearing to read "Trace McInturff".

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 5428.01  
Valid to May 31, 2025



*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*